

Specifications	Phosphate MW12 Phosphate	MW13 lodine	MW14 Iron
Resolution	0.01 ppm	0.1 ppm	0.01 ppm
Accuracy (@ 25 °C)	±0.04 ppm ±4% of reading	±0.1 ppm ±5% of reading	±0.04 ppm ±2% of reading
Typical EMC Dev.	±0.01 ppm	±0.1 ppm	±0.01 ppm
Light Source	Light Emitting Diode @ 525 nm	Light Emitting Diode @ 525 nm	Light Emitting Diode @ 525 nm
Light Detector	Silicon Photocell	Silicon Photocell	Silicon Photocell
Method	Adaptation of the Standard Methods fo the Examination of Water and Wastewater, 20th edition, Ascorbic Acid method. The reaction between phosphate and the reagent causes a blue tint in the sample.	Adaptation of the Standard Methods for the Examination of Water and Wastewater, 18th edition, DPD method. The reaction between iodine and the reagent causes a pink tint in the sample.	Adaptation of the EPA Phenantroline method 315B, for natural and treated waters. The reaction between iron and reagent causes an orange tint in the sample.
Environment	0 to 50°C (32 to 122 °F) max. 95% RH non-condensing	0 to 50°C (32 to 122 °F) max. 95% RH non-condensing	0 to 50°C (32 to 122 °F) max. 95% RH non-condensing
Battery Type	1 x 1.5V AAA	1 x 1.5V AAA	1 x 1.5V AAA
Auto-Shut Off	After 2 minutes of non-use	After 2 minutes of non-use	After 2 minutes of non-use
Dimensions	81.5 x 61 x 37.5 mm (3.2 x 2.4 x 1.5")	81.5 x 61 x 37.5 mm (3.2 x 2.4 x 1.5")	81.5 x 61 x 37.5 mm (3.2 x 2.4 x 1.5")
Weight	64 g (2.25 oz.)	64 g (2.25 oz.)	64 g (2.25 oz.)

Accessories

2720115 Phosphate powder reagent, (25 pcs)
2720316 Iodine powder reagent (25 pcs)
2720416 Iron powder reagent, (25 pcs)

/

Mi0011 Glass cuvets (2 pcs)
Mi0013 Stoppers for cuvets (2 pcs)

Ordering information:

All handy photometers are supplied in a carton box including 2 cuvets, 6 powder reagents, 1 x 1.5 V AAA battery and instructions.



FieldTech Solutions

Tel - +61 3 9676 9664 37 Wirraway Drive, Port Melbourne, VIC.3207 Email - sales@fieldtechsoln.com